

March 22, 2019

VIA ECFS

Marlene H. Dortch, Secretary
Federal Communications Commission
445 12th Street, SW
Room TWA325
Washington, DC 20554

**Re: Notice of *Ex Parte* Presentation
ET Docket No. 18-39**

Dear Ms. Dortch:

Sensible Medical Innovations Ltd. (“Sensible”), through its counsel, submits this letter to respond to comments by the GPS Innovation Alliance (“GPSIA”), which involves Sensible’s request for a limited waiver of the FCC’s Part 15 rules governing ultra-wideband medical imaging devices. Sensible seeks these waivers to market a medical device that uses dielectric sensing to provide non-invasive measurement of lung fluid (the “ReDS System”).¹

Following the comments from GPSIA, and in order to mitigate concerns of interference, Sensible proposes to modify the third frequency in its frequency table from 1164.0625 to 1163.0625 MHz,² avoiding the frequency range to which Section 15.513(e) applies.³

Respectfully submitted,

/s/ Michele C. Farquhar

Michele C. Farquhar
Tom Peters
J. Ryan Thompson
HOGAN LOVELLS US LLP
555 Thirteenth Street, NW
Washington, DC 20004
Phone: (202) 637-5663

Counsel to Sensible Medical Innovations Ltd.

Nadav Mizrahi
SENSIBLE MEDICAL INNOVATIONS LTD.
6 Meir Ariel Street
Netanya 4059300
Israel
POB 8702

cc: Hugh Van Tuyt
Jamison Prime

¹ Sensible, Request for Waiver, ET Docket No. 18-39 (filed Jan. 16, 2018); *Office of Engineering and Technology Seeks Comment on Sensible Medical Innovations Ltd’s Request for Waiver of Part 15 Ultra-Wideband Rules for Medical Imaging System*, Public Notice, 33 FCC Rcd 1307 (OET 2018).

² See Letter from Michele C. Farquhar, Counsel, Sensible, to Marlene H. Dortch, Secretary, FCC, ET Docket No. 18-39, at 2 (filed Sept. 5, 2018) (providing the full 16 frequency table).

³ 47 C.F.R. § 15.513(e).