FISH.
FISH & RICHARDSON

March 2, 2016

Terry G. Mahn

VIA ELECTRONIC FILING
Marlene H. Dortch, Secretary
Federal Communications Commission
445 12th Street, S.W.
Washington, DC 20554

RE: Kyma Medical Technologies Ltd (Kyma) Waiver of Part 15 of the Commission's Rules Applicable to Ultra-Wideband Devices, ET Docket No. 15-119:
NOTICE OF EX PARTE PRESENTATION

Dear Ms. Dortch:

I am writing pursuant to Section 1.1206(b) of the Commission's Rules to report that on March 1, 2016, I (on behalf of Kyma) spoke with Mark Settle, Chief, Policy and Rules Division, Office of Engineering and Technology, regarding the above-referenced proceeding. In response to Mr. Settle's request, I clarified that the sweep rate of Kyma's ultra-wideband medical imaging system ("Device") is 20 msec (not 10 msec) which means that the Device visits each relevant frequency once every 20 msec. I also clarified that the actual time to sweep over all frequencies is only 10 msec, but the Device then sleeps for 10 msec before starting the next sweep.

Pursuant to Sections 1.1206(b) and 1.49(f) of the Commission's Rules, this letter is being filed electronically with the Commission via the Electronic Comment Filing System. Should you have any questions regarding this letter, please contact the undersigned.

Respectfully submitted,

/s/ Terry Mahn

Terry Mahn
Outside Counsel for
Kyma Medical Technologies Ltd

cc Mark Settle