

1
2
3
4
5

**Before the
Federal Communications Commission
Washington, D.C. 20554**

In the Matter of)
)
)
Reassessment of Federal Communications) ET Docket No. 13-84
Commission Radiofrequency Exposure)
Limits and Policies)
)
)
Proposed Changes in the Commission's) ET Docket No. 03-137
Rules Regarding Human Exposure to)
Radiofrequency Electromagnetic Fields)
)

6
7
8


**Comments of the Cardiac Rhythm Management Device Committee (CRMD) / Working Group WG02 on
EMC Protocols of the Association for Advancement of Medical Instrumentation (AAMI)**

9
10

***This document solely represents the views of the AAMI CRMD / WG02 membership, and does not
necessarily represent a position of AAMI or FDA.***

11 Submitted by WG02 co-chairs:

12 Mr. Ronald Reitan, M.S.
13 Principal Electrical Engineer
14 Boston Scientific Corporation
15 4100 Hamline Avenue North
16 St. Paul, MN 55112
17 651-582-4529
18 Ronald.reitan@bsci.com



19
20
21 Mr. Seth J. Seidman, M.S.
22 Research Electrical Engineer
23 Center for Devices and Radiological Health
24 Food and Drug Administration
25 10903 New Hampshire Ave
26 White Oak Bldg 62 Room 1134
27 Silver Spring, MD 20993
28 (301) 796 – 2477
29 Seth.Seidman@fda.hhs.gov

30
31 Submitted: September 3, 2013
32

33 **INTRODUCTION**

34 The Association for Advancement of Medical Instrumentation (AAMI) CRMD committee is responsible
35 for development of a number of performance standards for implantable devices used to treat cardiac
36 arrhythmia and heart failure. The committee also serves as the US Technical Advisory Group to
37 International Standards Organization (ISO) TC150/SC6/JWG1, which is responsible for development of
38 international standards for active implantable cardiac devices. These committees worked closely in the
39 development of ISO 14117:2012, which is the guiding electromagnetic compatibility standard for all
40 implantable cardiac devices. This standard is currently being revised under the auspices of AAMI by a
41 working group WG02 reporting to the standing AAMI CRMD committee. Membership of the working
42 group consists of experts from all cardiac device manufacturers, allied manufacturers, independent
43 experts, and the FDA.

44

45 **SUMMARY**

46 We recommend that the Commission carefully consider its responsibility to “protect the public without
47 imposing an undue burden on industry”, and adopt harmonized human exposure levels as low as
48 deemed possible thereby limiting the impact upon those persons having active implantable cardiac
49 devices, in addition to the potential impact for biological effects for all patients who might encounter
50 these fields.

51

52 Given the critical nature of the 0-100 kHz frequency band in question, we strongly recommend that the
53 Commission consider development of either rules or guidelines for emitter manufacturers to which the
54 low frequency spectrum applies.

55

56 We further recommend that such guidance be developed collaboratively in conjunction with the AAMI
57 CRMD / WG02, as they possess considerable knowledge concerning the effects of low frequency
58 emissions upon implantable cardiac devices as well as potential risk mitigation methods that
59 manufacturers might apply. We ask the Commission to establish an ongoing liaison role with the AAMI
60 CRMD / WG02.

61

62 We urge the Commission to reconsider the need for rulemaking in regards to SAR exposure for
63 conductive implants. Should the Commission still feel there is a need to do so, we strongly recommend
64 the Commission to establish rules or guidance on this topic only after collaboration with the ISO
65 TC150/SC6/JWG 2 committee (author of ISO TS 10974). We ask the Commission to establish an ongoing
66 liaison role with the JWG 2.

67

68

69 **DETAILS**

70

71 Pursuant to Section 1.405 of the Federal Communications Commission’s (FCC’s) Rules, the AAMI CRMD
72 Committee /Working Group WG02 hereby submits comments in response to the *Further Notice of*
73 *Proposed Rule Making* (FNPRM) and *Notice of Inquiry* (NOI) issued by the Commission in the above-
74 captioned proceeding.

75

76 I. EXPOSURE LIMITS (Regarding paragraph 207-209 of the NOI)

77

78 The EMC standard ISO 14117:2012 “Active implantable medical devices — Electromagnetic
79 compatibility — EMC test protocols for implantable cardiac pacemakers, implantable cardioverter
80 defibrillators, and cardiac resynchronization devices” [1] establishes immunity requirements for

81 cardiac devices when patients are exposed to EM fields from 0 Hz to 3 GHz. This spectrum is divided
82 into several bands, with different testing approaches provided for each band, appropriate to the
83 type of fields and field absorption characteristic in the human body. To establish the immunity
84 levels, the authors of ISO 14117 and its predecessors (e.g., AAMI PC69 [2], ISO 14708-2 [3], ISO
85 14708-6 [4]) utilized information from a variety of known emitters, as well as the reference levels
86 established by ICNIRP 1998[5]. The MPE (reference) values from ICNIRP 1998 were multiplied by a
87 safety factor and used to predict the level at which interference might be expected at the sensing
88 ports of cardiac devices. These levels form the basis of conformance tests that have been in place
89 since the year 2000. In that time, hundreds of thousands of patients have received devices tested to
90 these limits, which are implied within the existing standard to provide a level of safety for patients
91 as they go about their lives in the general public environment.

92
93 Any standard or rulemaking that concerns MPE, whether in the US or elsewhere, is of importance to
94 the authors of ISO 14117, and ultimately to the patients whose safety is assured through conformity
95 with it. By themselves, MPE levels are not the concern of the authors of ISO 14117. However, if the
96 MPE levels are in turn used by standards organizations or rule-making bodies to set the allowable
97 emissions of equipment subject to such rules, then it may reasonably be expected that at some
98 point in the future, there will be a higher level of fields in either the general public or occupational
99 environments.

100
101 At this point two issues arise. First, the safety of patients having implanted devices tested to the
102 current device standards may no longer be assured. Secondly, to ensure safety in the future without
103 imposing undue cautions or warnings, (and attendant reduction in Quality of Life or QoL), devices
104 would need to meet a higher immunity standard. Unfortunately, due to the very nature of cardiac
105 devices, the signals they are designed to sense, and the desire to implement more features in a
106 smaller device size, device design for higher immunity levels, especially at frequencies below 30
107 MHz is a difficult challenge. These implantable devices are designed to sense low amplitude
108 (microvolts-millivolts) intrinsic cardiac signals in the frequency range from 0-500Hz. Accurate
109 sensing of these cardiac signals is critical for the appropriate operation of these devices and patient
110 safety.

111 It is therefore more likely that the sole outcome of an environment posing an increased interference
112 potential will be that manufacturers will have to provide considerably more cautionary and warning
113 language to their patients. Such an outcome leads to deterioration in patient QoL due to anxiety and
114 restriction of movement in their daily lives that is as deleterious as their underlying heart disease.

115
116 In regards to the Commission's question as to whether the exposure levels should be more
117 restrictive, less restrictive, or remain the same, our working group has no comment for frequencies
118 above 300 kHz where harmonized levels have already been established. We do however strongly
119 support the concept that the Commission harmonizes its MPE levels with those specified within
120 either the IEEE C95.1:2005[8] standard, or the ICNIRP 1998 Guidelines. Such harmonization will
121 allow our working group to revise its immunity requirements (for frequencies above 300 kHz) with
122 greater confidence that they will apply with equal efficacy on a global scale.

123
124 Clearly, there is a tradeoff here between patient safety / QoL and the needs of the non-medical
125 manufacturing and general public. We recommend that the Commission carefully consider its
126 responsibility to "protect the public without imposing an undue burden on industry", and adopt
127 harmonized MPE levels as low as deemed possible under this tradeoff.

128

129 We further recommend that the Commission establishes a liaison level of participation with the
130 AAMI CRMD / WG02 in order to determine ways to mitigate the problems outlined above. The
131 Commission could provide guidance, to be developed in conjunction with AAMI CRMD / WG02,
132 which would instruct equipment manufacturers as to possible design or use mitigations. Examples of
133 such mitigations include avoidance of amplitude modulation, choosing burst rates that are well
134 above 5 Hz thereby avoiding mimicry of cardiac activity, control over the exposure duration (either
135 by design, labeling or signage), and shielding. These mitigations can significantly reduce the safety
136 risk to persons having cardiac implants, and our working group would welcome the opportunity to
137 collaboratively develop with the Commission such guidance.

138 II. FREQUENCY RANGE (paragraph 229 of the NOI)

139
140
141 As previously stated, implantable cardiac devices perform critical physiological sensing of heart
142 rhythm in the frequency range of 0 to 500Hz, at voltages between 10 μ V and a few mV. This
143 bandwidth is required in order to capture signal components that represent all periods and types of
144 cardiac activity. The devices are generally designed to ignore signals above 1 kHz to the maximum
145 extent possible. Due to their miniature design, these devices rely upon digital filtering to accomplish
146 out of band rejection at low frequencies (up to 30 MHz), above which analog filtering becomes
147 feasible and necessary. For frequencies between 0 and 100 kHz, cardiac devices are extremely
148 vulnerable to emitters with high magnetic fields, in particular sources that employ pulse or duty
149 cycle modulation at low frequencies. Given the critical nature of the frequency band in question, we
150 recommend that the Commission should consider development of either rules or guidelines for
151 emitter manufacturers to which the low frequency spectrum applies.

152
153 We understand that there are no current restrictions in this frequency range. Therefore, it is
154 entirely possible that equipment is brought to the market that is FCC compliant, while at the same
155 time presenting a risk to persons with cardiac devices. Once such risks become reality, medical
156 device manufacturers must then react to protect the safety of their patients. This reaction can lead
157 to the aforementioned additional warnings to patients, incorporation of design changes in future
158 implantable devices, and collaboration with the equipment manufacturers. Collaboration with
159 equipment manufacturers is on a voluntary basis and usually long after the product has reached the
160 marketplace, by which time it is difficult to effect a change. By providing guidance to equipment
161 manufacturers, the Commission can reverse this state of affairs from one of reactive to proactive
162 collaboration between medical device manufacturers and manufacturers of potentially harmful
163 emitters.

164
165 We recommend that these rules or guidelines be developed collaboratively in conjunction with the
166 AAMI CRMD / WG02, as we possess considerable knowledge concerning the effects of low
167 frequency emissions upon implantable cardiac devices as well as potential risk mitigation methods
168 that manufacturers might apply. The FDA, medical device manufacturers, and manufacturers of
169 potentially harmful emitters must equally engage with the FCC to develop low frequency limits that
170 are in-line with physiological sensing constraints.

171 III. CONDUCTIVE IMPLANTED OBJECTS (paragraph 230 of the NOI)

172
173
174 The Commission has posed two questions in regards to conductive implanted objects, the first being
175 whether present volume-averaged SAR limits are protective for the more localized SAR that may

176 occur near the tip of a conductive object such as an implanted wire. Secondly, the Commission asks
177 whether “high levels of RF exposure may cause internal injury at the site of conductive implants”.

178
179 Our response to the latter question is clearly yes, as has been documented in several cases where
180 patients with implanted devices (and leads) were exposed and subsequently injured due to the high
181 RF fields associated with MRI systems. These cardiac devices had not specifically been designed and
182 labeled as conditionally approved for MR exposure. In the literature, during MRI, there are
183 examples of pacemakers not designed for the MR environment having inhibited (immediately life
184 threatening for a pacemaker dependent patient), devices have been permanently damaged, devices
185 have been reset, and distal electrodes have overheated. Inhibition occurs when the pacemaker
186 detects EMI as a normal cardiac rhythm and then shuts off (to avoid rate competition and also to
187 save battery life).

188
189 As to the former question on the protective nature of present SAR limits, exposure to other sources
190 of high levels of RF energy is a very complex question that, fortunately, has been addressed by other
191 standards development organizations and groups. For example, IEC 60601-2-33 [9] provides SAR
192 values that were derived in a manner such that localized temperatures will not result in tissue
193 damage. Also, ASTM standard F2182-11[10] governs the measurement of heating (e.g., orthopedic
194 devices). Finally, ISO TS 10974:2012 [11], sets forth information for the compatibility between MRI
195 exposure and active implantable medical devices.

196
197 These standards documents have all been developed in response to the needs of the medical device
198 regulatory agencies, including the FDA.

199 We urge the Commission to reconsider the need for rulemaking in regards to SAR exposure for
200 conductive implants. Should the Commission still feel there is a need to do so, we strongly
201 recommend the Commission establish rules or guidance on this topic only after collaboration with
202 the ISO TC150/SC6/JWG 2 committee (author of ISO TS 10974). There should be an ongoing liaison
203 role between the Commission and the JWG 2, composed of over 100 experts in the area of
204 interaction between conductive implants and modeling and prediction of the effects due to RF
205 exposure.

206 207 IV. TRANSMITTER POWER EXEMPTION THRESHOLD (paragraphs 121-126 of the FNPRM)

208
209 Implantable cardiac devices almost universally are equipped with telemetry radios. However, due to
210 the inherent need to provide maximum battery life, such radios are intentionally designed to have
211 low output power and limited range. We support the Commissions’ establishment of a 1 mW
212 exemption level in that it would simplify regulatory approval steps for all device manufacturers. We
213 further recommend however that the Commission consider increasing the exemption level to the
214 highest possible value consistent with SAR limits based upon IEEE standard and ICNIRP guidelines, as
215 long as MPE levels are consistent with internationally adopted standards for cardiac device
216 immunity (e.g., ISO 14117). This additional margin may in the future allow cardiac devices to be
217 exempted at higher power levels if necessary, either for enhanced range, data rate, or where a
218 rechargeable power source is available. This will benefit not only manufacturers with reduced
219 regulatory burden, but also patients who may enjoy higher levels of functionality.

220 221 V. REFERENCES

222 [1] ISO 14117:2012 Active implantable medical devices — Electromagnetic compatibility —
223 EMC test protocols for implantable cardiac pacemakers, implantable cardioverter
224 defibrillators, and cardiac resynchronization devices
225 [2] AAMI PC69:2007 Active implantable medical devices—Electromagnetic compatibility—
226 EMC test protocols for implantable cardiac pacemakers and implantable cardioverter
227 defibrillators
228 [3] ISO 14708-2:2005 Implants for surgery — Active implantable medical devices — Part 2:
229 Cardiac pacemakers
230 [4] ISO 14708-6:2010 Implants for surgery — Active implantable medical devices — Part 6:
231 Particular requirements for active implantable medical devices intended to treat
232 tachyarrhythmia (including implantable defibrillators)
233 [5] ICNIRP Guidelines for Limiting Exposure to Time-Varying Electric, Magnetic, and
234 Electromagnetic fields (up to 300 GHz). Health Physics, 74(4): 494–522, 1998.
235 [6] ICNIRP GUIDELINES FOR LIMITING EXPOSURE TO TIME-VARYING ELECTRIC AND
236 MAGNETIC FIELDS (1 HZ – 100 KHZ) PUBLISHED IN: HEALTH PHYSICS 99(6):818-836; 2010
237 [7] IEEE C95.6:2002 - IEEE Standard for Safety Levels With Respect to Human Exposure to
238 Electromagnetic Fields, 0-3 kHz
239 [8] IEEE C95.1:2005, IEEE Standard for Safety Levels with Respect to Human Exposure to
240 Radio-Frequency Electromagnetic Fields, 3 kHz to 300 GHz
241 [9] IEC 60601-2-33:2010 Medical Electrical Equipment - Part 2-33: Particular Requirements
242 for the Basic Safety and Essential Performance of Magnetic Resonance Equipment for
243 Medical Diagnosis
244 [10] ASTM F2182-11:2011 Standard Test Method for Measurement of Radio Frequency
245 Induced Heating Near Passive Implants During Magnetic Resonance Imaging
246 [11] ISO TS 10974:2012 Assessment of the safety of magnetic resonance imaging for
247 patients with an active implantable medical device